

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Data output equipment
ProfINET module

Name and address of the applicant

**LOVATO ELECTRIC S.p.A.**
Via Don Emilio Mazza 12
24020 Gorle (BG)
ITALY

Name and address of the manufacturer

LOVATO ELECTRIC S.p.A.
Via Don Emilio Mazza 12, 24020 Gorle (BG), ITALY

Name and address of the factory

LOVATO ELECTRIC S.p.A.
Via Don Emilio Mazza 12, 24020 Gorle (BG), ITALY

Ratings and principal characteristics

Input voltage: 5 V d.c.
Overvoltage Category III
Class of protection II

Trademark / Brand (if any)

Lovato electric

Model/type Ref.

EXP1019

Additional information (if necessary)

Conditions of Acceptability: The equipment shall be supplied from a source providing separation from hazardous live parts by means of double or reinforced insulation.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010
IEC 61010-1:2010/AMD1:2016as shown in the Test Report Ref. No.
which forms part of this certificate

ELS-V200220020060A-000

Ratings and principal characteristics (continued)



(Ralph Fischer)